



UNITED STATES PATENT AND TRADEMARK OFFICE

A

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/747,558	12/23/2000	Mehmet C. Oz	20076-51	5175

7590 03/30/2005
William H Doppert
Reed Smith LLP
599 Lexington Avenue
29th Floor
New York, NY 10022

EXAMINER

WOO, JULIAN W

ART UNIT	PAPER NUMBER
----------	--------------

3731

DATE MAILED: 03/30/2005

Please find below and/or attached an Office communication concerning this application or proceeding.

81

Office Action Summary	Application No.	Applicant(s)	
	09/747,558	OZ ET AL.	
	Examiner	Art Unit	
	Julian W. Woo	3731	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 13 January 2005.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 37,38,40-49 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 37,38,40-49 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some * c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- | | |
|--|---|
| 1) <input type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input type="checkbox"/> Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____ | 6) <input type="checkbox"/> Other: _____ |

DETAILED ACTION

Claim Rejections - 35 USC § 102

1. The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

2. Claims 37, 38, 40, 42-44, and 46-48 are rejected under 35 U.S.C. 102(e) as being anticipated by Yoon (5,620,452). Yoon discloses, in figure 11 and in col. 1, lines 22-58 and col. 3, lines 6-40, a fastener having a first plate (18), a pair of laterally-spaced apart penetrating elements (50) on the first plate, a second plate (16), a pair of holes (36) on the second plate for receiving the penetrating elements, and a hinge (at 14), where the first plate is free of additional penetrating elements spaced from the pair of laterally-spaced penetrating elements, where the penetrating elements pass entirely through the holes, where the fastener is composed of a biocompatible, non-immunogenic material (e.g., titanium), and where the fastener is adapted to be introduced via a percutaneous, endoscopic approach. Note: The introductory statement of intended use ("for securing coapted valve leaflets together") has been carefully considered but deemed not to impose any structural limitations on the claims patentably distinguishable over Yoon's device, which is capable of being used as claimed if one desires to do so.

Art Unit: 3731

Claim Rejections - 35 USC § 103

3. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

4. Claims 41, 45, and 49 are rejected under 35 U.S.C. 103(a) as being unpatentable over Yoon. Yoon discloses the invention substantially as claimed, but does not specify the plate dimensions as claimed. Nevertheless, it would have been a matter of design choice to dimension the plates in the ranges of dimensions as claimed. The choices would be dependent upon the desired size for the fastener and upon the place of application for the fastener.

Response to Arguments

5. Applicant's arguments filed on January 13, 2005 have been fully considered but they are not persuasive. See the rejection above. Moreover, Yoon teaches, in col. 6, lines 28-31, that "the number of tissue penetrating elements per row can be varied

Art Unit: 3731

according [to] the size of the clip and the vessel or organ to be occluded"--even a single pair of penetrating elements. Also, the orientation of the pair(s) of penetrating elements with respect to the plates or "arms" can be varied according to the desired degrees of occlusion (e.g., see figures 16 and 17).

Conclusion

6. **THIS ACTION IS MADE FINAL.** Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

7. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Julian W. Woo whose telephone number is (571) 272-4707. The examiner can normally be reached Mon.-Fri., 7:00 AM to 3:00 PM Eastern Time, alternate Fridays off.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Anh Tuan Nguyen can be reached on (571) 272-4963. The fax phone

Art Unit: 3731

number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

A handwritten signature in black ink, reading "Julian W. Woo". The signature is fluid and cursive, with the first name "Julian" being more prominent than the last name "Woo".

Julian W. Woo
Primary Examiner

March 29, 2005